



Application Information / Information de soumission	
*Licence Name/Nom de l'homologation: MAGNETIQ VIRAL RNA EXTRACTION KIT	Application Number/Numéro de soumission 314136
Licence # / # de l'homologation: 0	Application Type/Type de soumission: Interim Order
Manufacturer/Fabricant: GALENVS SCIENCES INC.	Company ID/Identificateur d'entreprise: 153721
Risk Class: 1	Rationale: Class I by IVDD Rule 8
*Licence Type/Type d'homologation: Test Kit	Rationale:
<input type="checkbox"/> Contains Controlled Substance(s) <input type="checkbox"/> Contains Drug(s) <input type="checkbox"/> Contains Biological Material(s)	
*Intended Use and/or Indications for Use/ Utilisation Prévue et/ou Indications “The intended use for this device is self-evident to the intended user, therefore an intended use statement is not necessary for which the medical device is manufactured, sold or represented, as stated in Regulation 10(h) of the Interim order respecting the importation and sale of medical devices for use in relation to COVID-19”	

OEM Licence Information	
OEM Licence Name :	OEM Manufacturer :
OEM Intended Use and/or Indications for Use	

Reason for Change		Comment(s)
Change to classification of a device	<input type="checkbox"/>	
Manufacturer name change	<input type="checkbox"/>	
License name change	<input type="checkbox"/>	
Device name change	<input type="checkbox"/>	
Change to the purpose/indication of license	<input type="checkbox"/>	
Addition of device(s)	<input type="checkbox"/>	
Deletion of device(s)	<input type="checkbox"/>	

Reason for Change		Comment(s)
Change in name and/or address of the Private Label Manufacturer	<input type="checkbox"/>	
Private Label License name change	<input type="checkbox"/>	
Private Label Device name change	<input type="checkbox"/>	
Addition of device(s)	<input type="checkbox"/>	
Deletion of device(s)	<input type="checkbox"/>	

Certificate Screening Checklist:

Cert # (new):	Cert Revisions / Comments (If Applicable):
Cert. # (old):	



Replacing Existing Cert on File (Y/N):	
--	--

Criteria	conforms	Comments/info for MDS
Issued to full name of manufacturer as it appears on application/licence and label.	<input type="checkbox"/>	
Issued to complete civic address matching application/licence and label.	<input type="checkbox"/>	
Criteria are ISO 13485:2016 and Medical Devices Regulations – Part 1 – SOR 98/282	<input type="checkbox"/>	
Scope activities limited to design, development, manufacture, production, servicing, installation, or distribution.	<input type="checkbox"/>	
Activities include “manufacture” or “production” Activities include “design” or “development and development” for class III/IV devices.	<input type="checkbox"/>	
Scope is unambiguous and covers app./lic. devices. Does not contain product names/models/licence numbers.	<input type="checkbox"/>	
Auditing Organisation is Authorized or Recognized	<input type="checkbox"/>	
Statement of Authorization or Recognition.	<input type="checkbox"/>	
Field labelled “Effective Date”	<input type="checkbox"/>	
Field labelled “Expiry Date”	<input type="checkbox"/>	
Validity period ≤ 3 years	<input type="checkbox"/>	
Unique identification code labelled “certificate number” or “certification document number”	<input type="checkbox"/>	<input type="checkbox"/> new <input type="checkbox"/> revised
Name, title, and signature of certification authority	<input type="checkbox"/>	
Pagination (page x or y) included on all pages . All pages present.	<input type="checkbox"/>	
Method to verify validity	<input type="checkbox"/>	

Screening Decision	
<div style="border: 1px solid black; padding: 2px;"> Accept ▼ </div>	
<div style="border: 1px solid black; padding: 2px;"> Liem Whelan ▼ </div> <hr/> Device Licensing Services Division Medical Devices Bureau	Date: May 26 th , 2020



Check GMDN

GMDN Name	GMDN Code	GMDN Description	Issue	Action for Screener	Regulatory Action	Reference Material
pH-adjusted buffer solution IVD	42693	A pH-adjusted buffer solution intended to be used alone or in combination with other IVDs in the processing, staining, and/or testing of clinical laboratory specimens.				